NEVO+600M

MEDICAL DATA SHEET



AC/DC Modular Configurable PSU



The NEVO+600M modular configurable medical power supply is the smallest in its class and the ultimate solution for demanding medical applications where size, power density and weight matter. Its tiny footprint of 5" x 3" x 1.61" weighs only 600 grams and delivers an incredible 600 Watts - equating to a power density of 25 Watts per cubic inch. The input module can accommodate up to four isolated output modules which can be configured into a high power 5"x 3" single output power supply or a multiple output power supply with up to 8 isolated outputs. Standard features include intelligent fan control providing optimised airflow for various load and temperature conditions, wide output voltage adjust, parallel and series connection of modules and an isolated 5V 1A bias supply. A low noise fan option is available that allows you to use this innovative power supply in even the quietest of environments. The series is approved to latest medical standards and features market leading specifications and design in application support.

MAIN FFATURES

• 600 Watts in 5" x 3" x 1.61"	Efficiency up to 89%	 Up to 8 isolated outputs
 User and field configurable 	 Intelligent fan control 	 Low noise option (ML version)
 Wide output voltage adjust range 	 Parallel & series connection of modules 	 IEC/UL60601-1 Ed. 3 & -1-2 Ed. 4 (EMC)
Remote current & voltage programming	 Standard 5V 1A bias supply 	3 Year warranty
	 Accurate current sharing 	

APPLICATIONS

Medical & diagnostic equipment	 Telecommunications 	• Lasers
 Test & Measurement equipment 	 Laboratory & Analysis equipment 	LED lighting
 Robotics 	Display	 Retrofit of legacy PSUs
• Oil & Gas	Avionics	

CUSTOMER BENEFITS

Fast time to market	 Proven technology 	 Technology consolidation
 24 hrs samples from distribution 	 Eliminates custom design costs 	 Supplier consolidation
 Safety & EMC certified 	 Field replaceable 	 Redundant manufacturing sites
 World class engineering support 	 Low cost of ownership 	

DOC-DTS-002-09, NEVO+600M Medical Datasheet





INPUT MODULE SPECIFICATIONS						
Parameter	Details	Min	Typical	Max	Units	
AC Input Voltage	Nominal range is 100V _{RMS} to 240V _{RMS}	85		264	V_{RMS}	
AC Input Frequency	Contact factory for 400Hz operation.	47	50/60	63	Hz	
DC Input Voltage	Not covered by safety approvals. Contact Vox Power.	120		300	V_{DC}	
Output Power Rating	De-rate linearly from 600Watts at 120V _{RMS} to 450Watts at 85V _{RMS}			600	Watts	
Input Current	600Watts output at 120 V _{RMS} input			6	Amps	
Input Current Limit	Maintains power factor		8		Amps	
Inrush Current	265V _{RMS} , 25°C (cold start)			20	Amps	
Fusing	Live line fused (5x20 Fast acting)			8	Amps	
Efficiency	See graphs		86	89	%	
No load Power consumption	All outputs fitted and disabled/enabled 21/28			Watts		
Power Factor	Typical value for 300 Watts output at 240Vrms input		0.96	0.99		
Holdup	600Watts output at 120V _{RMS} input	17	20	21	mS	
UVP	Turn on under voltage protection	78		84	V_{RMS}	
Over temperature	Internally monitored.	115		125	°C	
Reliability (1)	Input module		•	1.207	FPMH	
	Fan			2.7	FPMH	
Warranty	Standard terms and conditions apply 3 Y					
Size	Size 133.7 (L) x 77.7 (W) x 41.0 (H). See diagram for tolerance details					
Weight	360 + 60 per output module				Grams	
Note 1.	30°C base & ambient, 100% load, SR332 Issue 2 Method I, Case 3, Ground, Fixed, Contro	lled	•			

GLOBAL SIGNALS SPECIFICATIONS						
Parameter	Details	Min	Typical	Max	Units	
Bias Voltage	Two isolated Bias Outputs available	4.8	5	5.2	Volts	
Bias Current	Hiccup type current limit	0		1	Amps	
AC_OK Voltage	Low output level High output level	0 3.5	0.2 4.5	1 5.2	Volts	
AC_OK Current		-10		20	mA	
Power Good Voltage	Low output level. internal 10kΩ pull down. High output level. PNP open collector.	0 8	0 10	0 15	Volts	
Power Good Current	Open collector output. Current source only. All Slots.			20	mA	
Global Inhibit Voltage	Low input level High input level	0 3		1 15	Volts	
Global Inhibit Current	5k input impedance.	0.6		3	mA	
Inhibit Voltage	Low input level. All slots. High input level. All slots.	0 2.5		1 15	Volts	
Inhibit Current	10k input impedance. All slots.	0.25		1.5	mA	

	OUTPUT MODULE SPECIFICATION SUMMARY											
MODEL	Out	put Volta	age	Output	Rated	Peak	Load	Line	Cross	Ripple &	FPMH (1)	Feature
MODEL	Min.	Nom.	Max.	Current	Power	Power	Reg.	Reg.	Reg.	Noise	FFIVIE '	Set (2)
OP1	1.5V	5V	7.5V	25A	125W	187.5W	±50mV	±5mV	±10mV	50mV _{PP}	0.5	ABCDEFG
OP2	4.5V	12V	15V	15A	150W	225W	±100mV	±12mV	±24mV	120mV _{PP}	0.5	ABCDEFG
OP3	9V	24V	30V	7.5A	150W	225W	±150mV	±24mV	±48mV	240mV_{PP}	0.5	ABCDEFG
OP4	18V	48V	58V	3.75A	150W	217.5W	±300mV	±48mV	±96mV	480mV _{PP}	0.5	ABCDEFG
OP5	3.3V	12V	15V	5A	2x 75W	2x 75W	±50mV	±12mV	±24mV	240mV_{PP}	0.75	AFG
OP8	23.2V	24V	24.7V	3.125A	2x 75W	2x 75W	±100mV	±24mV	±48mV	480mV _{PP}	0.75	AFG
OPA2 ⁽³⁾	4.5V	12V	15V	25A	300W	375W	±100mV	±12mV	±24mV	120mV _{PP}	0.5	ABCDEFGH
OPA3 ⁽³⁾	9V	24V	30V	15A	300W	450W	±150mV	±24mV	±48mV	240mV _{PP}	0.5	ABCDEFGH
Note 1.	Note 1. Output module, 30°C base, 100% load, SR332 issue 2 Method I, Case 3, Ground, Fixed, Controlled											
Note 2.	Note 2. A = Remote Sense, B = External Voltage control, C = External constant current control, D = Current output signal, E = Current share, F = Over Voltage protection,											
	G = Over temperature protection, H = Dual Slot module											

G = Over temperature protection, H = Dual Slot module

Note 3. Can only be used with NEVO+600 chassis with date codes from 2048 onwards. eg. 2048C080000 can use A2 or A3 module, 2047C089999 cannot use A2 or A3 module.

SAFETY SPECIFICATIONS					
Parameter	Details	Max	Units		
	Input to Output (2 MOPP). Do not perform test on assembled unit ⁽¹⁾	4000	V_{AC}		
Isolation Voltages	Input to Chassis (1 MOPP)	1500	V_{AC}		
	Global signals (J2) to Output/Chassis	250	V_{DC}		
	Output to Output/Chassis (Standard modules)	250	V_{DC}		
Earth Leakage Current	Normal condition, 264Vac, 63Hz, 25°C	300	uA		
Touch Leakage Current	Standard modules NC/SFC	20/200	uA		
Patient Leakage Current	Standard modules 264Vac, 63Hz, 25°C NC/SFC ⁽²⁾		uA		
Note 1. Testing an assembled	unit to 4000V _{AC} may cause damage. Please refer to application note (APN-002) on Vox Power web	osite or contact Vox Power repr	esentative.		
Note 2. Not Applicable					

INSTALLATION SPECIFICATIONS							
Parameter Details Parameter Details							
Equipment class	I	Flammability Rating	94V-2				
Overvoltage category	II	Ingress protection rating	IP10				
Material Group	IIIb (indoor use only)	ROHS compliance	2011/65/EU & 2015/863/EU				
Pollution degree	2	Intended usage environment	Home Healthcare				



ENVIRONMENTAL SPECIFICATIONS								
D	D - 1			Non-Op	erational	Opera	11	
Parameter	Details	-		Min	Max	Min	Max	- Units
Air Temperature	Operational limits subject to appropr	-40	+85	-20	70	°C		
Humidity	Relative, non-condensing			5	95	5	95	%
Altitude				-200	5000	-200	3000	m
Air Pressure				52	106	69	106	kPa
Noise Level	Variable. Measured 1m from fan inta	ke.		-	-	36	62	dBA
Shock	3000 bumps at 10G (16ms) half sine v							
Vibration	1.5G 10 to 200Hz sine wave, 20G for 1	15min in 3 axes random vibration						
	ELECTROMAGNETIC COMPLIANCE – E							
Phenomenon		Basic EMC Standard		Tes	t Details			
Radiated emissions	s, electric field	EN55011/22, FCC		Clas	s B compliant			
Conducted emission	ons	EN55011/22, FCC part 15, CISPR 22/11 Class B compliant						
Harmonic Distortio	n	IEC61000-3-2 Compliant						
Flicker & Fluctuatio	n	IEC61000-3-3	Compliant					
	ELE	CTROMAGNETIC COMPLIA	ANCE –	IMMUN	TY			
Phenomenon		Basic EMC Standard	Test D	Details				
Electrostatic discha	nrge	IEC61000-4-2	Test lev	rel 4: 15kV ai	r, 8kV contact			
Radiated RF EM fiel	lds	IEC61000-4-3	Test Lev	vel 3: (10V/n	n, 80MHz-2.7GI	Hz) sine wave	AM 80% 1kH	Z
Proximity fields from RF wireless communications		IEC61000-4-3						
equipment		IEC01000-4-3	rest iev	Test levels as per IEC60601-1-2:2014 Table 9				
Electrical Fast Trans	sients/bursts	IEC61000-4-4	Test Lev	Test Level 3: (2kV Power, 1kV I/O) 5kHz(ed3) & 100kHz(ed4)				
Surges		IEC61000-4-5	Test Lev	Test Level 3: 1kV L-N, 2kV L-E				
Conducted disturb	ances induced by RF fields	IEC61000-4-6	Test Lev	Test Level 3: 10V, 0.15 to 80Mhz sine wave AM 80% 1kHz				

Criterion A = No degradation of performance or loss of function.
 Criterion B = Temporary degradation of performance or loss of function is allowed, provided the function is self-recoverable.
 Criterion C = Temporary loss of function is allowed but requires operator intervention to recover.
 Tested at nominal range (100V to 240V). Line deratings applied where appropriate.

IEC61000-4-11& SEMI-F47-0706 (2)

IEC61000-4-8

IEC61000-4-11


AGENCY APPROVALS					
Standard	Details	File			
IEC 60601-1:2005 + CORR1 2006 + CORR2: 2007 + A1:2012	Medical electrical equipment Part 1: General requirements for basic safety and essential performance	UL: E316486			
EN60601-1:2006 + A11:2011 + A1:2013 + A12:2014	Medical electrical equipment Part 1: General requirements for basic safety and essential performance				
CAN/CSA-C22.2 No. 60601-1 (2008)	Medical Electrical Equipment Part 1: General Requirements for Basic Safety and Essential Performance				
ANSI/AAMI ES60601-1 (2005 + C1:09 + A2:10)	Medical Electrical Equipment Part 1: General Requirements for Basic Safety and Essential Performance				
CE MARK	LVD 2014/35/EU, EMC 2014/30/EU				
CB certificate and report available on request					


Test level 4: 30A/m 50Hz

0% 10ms, 0% 20ms, 80% 1s, 80% 10s, 90% continuous (Criterion A)

70% 0.5s, 40% 0.2s (Criterion A at 240V and Criterion B at 100V)

0% 250/300 cycle as per IEC60601-1-2:2014 (Criterion B)






Power Frequency Magnetic Fields

Voltage Dips

Voltage interruptions

EA-0193







DOC-DTS-002-09, NEVO+600M Medical Datasheet

EA-0193







All specifications are believed to be correct at time of publishing. Vox Power Ltd reserves the right to make changes to any of its products and to change or improve any part of the specification, electrical or mechanical design or manufacturing process without notice. Vox Power Ltd does not assume any liability arising out of the use or application of any of its products and of any information to the maximum extent permitted by law. No license, express or implied, by estoppel or otherwise, to any intellectual property rights is granted by this document or by any products of Vox Power Ltd. VOX POWER LTD DISCLAIMS ALL WARRANTIES AND REPRESENTATIONS OF ANY KIND WHETHER EXPRESS OR IMPLIED, INCLUDING, BUT NOT LIMITED TO, IMPLIED WARRANTIES OF SUITABILITY, FITNESS FOR PURPOSE, MERCHANTABILITY AND NONINFRINGEMENT.

Please consult your local distributor or Vox Power directly to ensure that you have the latest revision before using the product and refer to the latest relevant user manual for further information relating to the use of the product. Vox Power Ltd products are not intended for use in connection with life support systems, human implantations, nuclear facilities or systems, aircraft, spacecraft, military or naval missile, ground support or control equipment used for the purpose of guidance navigation or direction of any aircraft, spacecraft or military or naval missile or any other application where product failure could lead to loss of life or catastrophic property damage. The user will hold Vox Power Ltd harmless from any loss, cost or damage resulting from its breach of these provisions.

DOC-DTS-002-09, NEVO+600M Medical Datasheet